



Confirmation of Product Type Approval

Company Name: CLASS INSTRUMENTATION LTD

Address: 837 GARRATT LANE, LONDON, United Kingdom, SW17 0PG

Product: Ultrasonic Level Measurement System, Level Sensor/Transmitter

Model(s): PLI Digital

Endorsements:

Certificate Type	Certificate Number	Issue Date	Expiry Date
Product Design Assessment (PDA)	24-0011328-PDA	18-JUL-2024	01-JUL 2029
Manufacturing Assessment (MA)	NA		NA
Product Quality Assurance (PQA)	NA		NA

Tier

2 - Type Approved, unit certification not required

Intended Service

Ultrasonic portable liquid level gauge for use on ABS Classed Vessels.

Description

Ultrasonic portable liquid level gauge for determination of the liquid level of CO₂, Halon, FM200, Propane and any liquefied gas under pressure.

Ratings

Power Supply: 4 x 1.5v batteries

Service Restrictions

Unit Certification is not required for this product. If the manufacturer or purchaser request an ABS Certificate for compliance with a specification or standard, the specification or standard, including inspection standards and tolerances, must be clearly defined.

Comments

The Manufacturer has provided a declaration about the control of, or the lack of Asbestos in this product.

Operators are to be suitably trained in the operation of the equipment.

Notes, Drawings and Documentation

Drawing No. Correspondence, PDA Application Form - Revalidation 14-LD1105851-PDA, Revision: -, Pages: 1